



*Manufacturer
of toxic gas
detection systems*

**BP-24V Battery Pack—used with GF1900
Sulfuryl fluoride [Vikane[®] / ProFume[®]]
Monitor**



SPECIFICATIONS

Power Source	Two 12V rechargeable gel-type Pb/PbO ₂ batteries
Maximum instrument running time provided by battery pack	75 minutes
Interconnect cable to instrument	5 ft (1.52m); MS-type connector
Carrying Case	Cordura [®] /Nylon, provided with 4 ft (1.22m) adjustable shoulder strap
Dimensions	6.5" H x 8.5" W x 5" D (165 x 216 x 127 mm)
Weight	13 lb (5.9 kg)

[Material Safety Data Sheet for Battery Pack](#)



interscan corporation

P.O. Box 2496 • Chatsworth, CA 91313-2496

1 800 458-6153 (US & Canada) (818) 882-2331 FAX (818) 341-0642

Copyright ©1996-2008 Interscan Corporation. All rights reserved. All other trademarks are the property of their respective owners.